

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060519\
 Data File : VN056004.D
 Acq On : 5 Jun 2019 12:51
 Operator : JC/SP
 Sample : K3151-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AU-06-060419-A

Quant Time: Jun 06 01:46:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	374288	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	550600	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	517934	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	175455	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	174924	47.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.14%	
35) Dibromofluoromethane	7.59	113	159514	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
50) Toluene-d8	10.09	98	656788	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.06%	
62) 4-Bromofluorobenzene	12.40	95	214845	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
Target Compounds						
16) Acetone	3.82	43	16889	9.502	ug/l	96
20) Methylene Chloride	4.54	84	37022	10.110	ug/l	96
43) Isopropyl Acetate	8.17	43	11814	1.731	ug/l #	86
95) Naphthalene	15.13	128	4167	4.310	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060519\
Data File : VN056004.D
Acq On : 5 Jun 2019 12:51
Operator : JC/SP
Sample : K3151-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AU-06-060419-A

Quant Time: Jun 06 01:46:59 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

