

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060519\
 Data File : VN056007.D
 Acq On : 5 Jun 2019 14:06
 Operator : JC/SP
 Sample : K3163-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-02-060419-A

Quant Time: Jun 06 01:56:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	392032	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	587153	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	558995	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	192719	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	184987	47.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.04%	
35) Dibromofluoromethane	7.59	113	168109	46.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.34%	
50) Toluene-d8	10.09	98	705067	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
62) 4-Bromofluorobenzene	12.40	95	232151	49.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.50%	
Target Compounds						
16) Acetone	3.82	43	34488	18.525	ug/l	98
20) Methylene Chloride	4.54	84	69526	18.127	ug/l	96
43) Isopropyl Acetate	8.17	43	14615	2.008	ug/l #	90
52) Toluene	10.16	92	14532	1.523	ug/l	98
95) Naphthalene	15.13	128	9800	4.833	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060519\
Data File : VN056007.D
Acq On : 5 Jun 2019 14:06
Operator : JC/SP
Sample : K3163-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-02-060419-A

Quant Time: Jun 06 01:56:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

