

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060519\
 Data File : VN056009.D
 Acq On : 5 Jun 2019 14:56
 Operator : JC/SP
 Sample : K3150-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IRONBOUND-TP

Quant Time: Jun 06 02:02:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	379682	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	569629	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	539328	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	189582	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	183083	48.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.12%	
35) Dibromofluoromethane	7.59	113	163616	46.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.64%	
50) Toluene-d8	10.09	98	675028	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
62) 4-Bromofluorobenzene	12.40	95	224685	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
Target Compounds						
16) Acetone	3.82	43	25460	14.121	ug/l	97
20) Methylene Chloride	4.55	84	16405	4.416	ug/l	88
43) Isopropyl Acetate	8.17	43	14106	1.997	ug/l #	90
95) Naphthalene	15.13	128	20007	5.888	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060519\
Data File : VN056009.D
Acq On : 5 Jun 2019 14:56
Operator : JC/SP
Sample : K3150-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IRONBOUND-TP

Quant Time: Jun 06 02:02:45 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

