

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060520\
 Data File : VN061663.D
 Acq On : 5 Jun 2020 12:47
 Operator : JC/MD
 Sample : L2904-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DAY-1

Quant Time: Jun 06 03:41:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	300883	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	524575	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	510391	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	239528	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	173338	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
35) Dibromofluoromethane	7.55	113	150240	56.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.92%	
50) Toluene-d8	10.06	98	621587	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
62) 4-Bromofluorobenzene	12.38	95	230009	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
Target Compounds						
16) Acetone	3.78	43	7622	1.552	ug/l	99
91) 1,2-Dichlorobenzene	13.63	146	16091	2.131	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060520\
Data File : VN061663.D
Acq On : 5 Jun 2020 12:47
Operator : JC/MD
Sample : L2904-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DAY-1

Quant Time: Jun 06 03:41:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

