

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060521\
 Data File : VN067223.D
 Acq On : 5 Jun 2021 11:22
 Operator : JC/MD
 Sample : VN0605WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0605WBL01

Quant Time: Jun 07 01:13:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 04 19:00:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	264642	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.973	114	455261	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	426953	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	182806	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	169352	46.946	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.900%
35) Dibromofluoromethane	8.027	113	137373	48.674	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.340%
50) Toluene-d8	10.446	98	614145	53.430	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.860%
62) 4-Bromofluorobenzene	12.736	95	186481	42.124	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.240%
Target Compounds						
95) Naphthalene	15.512	128	6171	5.116	ug/l	Qvalue # 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060521\
Data File : VN067223.D
Acq On : 5 Jun 2021 11:22
Operator : JC/MD
Sample : VN0605WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0605WBL01

Quant Time: Jun 07 01:13:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 04 19:00:20 2021
Response via : Initial Calibration

