

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060521\
 Data File : VN067226.D
 Acq On : 5 Jun 2021 13:41
 Operator : JC/MD
 Sample : M2589-29
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 18-B9(4-6)

Quant Time: Jun 07 01:14:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 04 19:00:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	264705	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.973	114	440421	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	422069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	184987	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	156310	43.320	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	86.640%
35) Dibromofluoromethane	8.027	113	132494	48.527	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.060%
50) Toluene-d8	10.446	98	572969	51.527	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.060%
62) 4-Bromofluorobenzene	12.736	95	181129	42.293	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.580%
Target Compounds						
					Qvalue	
16) Acetone	4.304	43	7522	6.042	ug/l	96
20) Methylene Chloride	5.106	84	6471	2.064	ug/l #	97
43) Isopropyl Acetate	8.568	43	7495	1.005	ug/l #	82
95) Naphthalene	15.509	128	6472	5.142	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060521\
Data File : VN067226.D
Acq On : 5 Jun 2021 13:41
Operator : JC/MD
Sample : M2589-29
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
18-B9(4-6)

Quant Time: Jun 07 01:14:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 04 19:00:20 2021
Response via : Initial Calibration

