

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060521\  
 Data File : VN067230.D  
 Acq On : 5 Jun 2021 15:15  
 Operator : JC/MD  
 Sample : M2589-33  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 18-B10(4-6)

Quant Time: Jun 07 01:15:38 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060421W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jun 04 19:00:20 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 8.091  | 168   | 219430   | 50.000   | ug/l  | # 0.00   |
| 34) 1,4-Difluorobenzene     | 8.973  | 114   | 381090   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 11.749 | 117   | 356053   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.680 | 152   | 158254   | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 8.442  | 65    | 140594   | 47.004   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =     | 94.000%  |
| 35) Dibromofluoromethane    | 8.027  | 113   | 114015   | 48.260   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =     | 96.520%  |
| 50) Toluene-d8              | 10.446 | 98    | 485169   | 50.424   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =     | 100.840% |
| 62) 4-Bromofluorobenzene    | 12.736 | 95    | 158374   | 42.738   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =     | 85.480%  |
| Target Compounds            |        |       |          |          |       |          |
|                             |        |       |          |          |       | Qvalue   |
| 16) Acetone                 | 4.296  | 43    | 16601    | 16.085   | ug/l  | 91       |
| 20) Methylene Chloride      | 5.101  | 84    | 6188     | 2.381    | ug/l  | 93       |
| 95) Naphthalene             | 15.509 | 128   | 20575    | 7.090    | ug/l  | 99       |
| -----                       |        |       |          |          |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060521\  
Data File : VN067230.D  
Acq On : 5 Jun 2021 15:15  
Operator : JC/MD  
Sample : M2589-33  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
18-B10(4-6)

Quant Time: Jun 07 01:15:38 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060421W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jun 04 19:00:20 2021  
Response via : Initial Calibration

