

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060521\
 Data File : VN067232.D
 Acq On : 5 Jun 2021 16:02
 Operator : JC/MD
 Sample : M2589-35
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 18-B11(2-4)

Quant Time: Jun 07 01:16:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 04 19:00:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	217715	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.973	114	382360	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	379827	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	164118	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	141234	47.590	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.180%		
35) Dibromofluoromethane	8.027	113	115717	48.818	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.640%		
50) Toluene-d8	10.446	98	500925	51.889	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	103.780%		
62) 4-Bromofluorobenzene	12.736	95	164360	44.206	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.420%		
Target Compounds					Qvalue	
16) Acetone	4.299	43	13991	13.663	ug/l	97
20) Methylene Chloride	5.109	84	5848	2.268	ug/l #	93
43) Isopropyl Acetate	8.566	43	15551	2.402	ug/l #	82
95) Naphthalene	15.512	128	17424	6.601	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060521\
Data File : VN067232.D
Acq On : 5 Jun 2021 16:02
Operator : JC/MD
Sample : M2589-35
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
18-B11(2-4)

Quant Time: Jun 07 01:16:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 04 19:00:20 2021
Response via : Initial Calibration

