

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060521\
 Data File : VN067233.D
 Acq On : 5 Jun 2021 16:25
 Operator : JC/MD
 Sample : M2589-36
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 18-B11(4-6)

Quant Time: Jun 07 01:16:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 04 19:00:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	204816	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.973	114	363246	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	362145	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.680	152	158598	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	135970	48.702	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.400%
35) Dibromofluoromethane	8.027	113	109497	48.624	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.240%
50) Toluene-d8	10.446	98	478884	52.216	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.440%
62) 4-Bromofluorobenzene	12.736	95	156188	44.218	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.440%
Target Compounds						
					Qvalue	
16) Acetone	4.304	43	21979	22.815	ug/l	96
20) Methylene Chloride	5.106	84	5944	2.450	ug/l	90
43) Isopropyl Acetate	8.568	43	15960	2.595	ug/l #	83
95) Naphthalene	15.509	128	29380	8.222	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060521\
Data File : VN067233.D
Acq On : 5 Jun 2021 16:25
Operator : JC/MD
Sample : M2589-36
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
18-B11(4-6)

Quant Time: Jun 07 01:16:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 04 19:00:20 2021
Response via : Initial Calibration

