

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060521\
 Data File : VN067234.D
 Acq On : 5 Jun 2021 16:49
 Operator : JC/MD
 Sample : M2589-37
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FOUNDATION-EXCAVATION-SAMP

Quant Time: Jun 07 01:16:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 04 19:00:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	212456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.973	114	376032	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	377981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	163573	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	140390	48.476	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.960%		
35) Dibromofluoromethane	8.027	113	112221	48.140	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.280%		
50) Toluene-d8	10.448	98	499881	52.652	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	105.300%		
62) 4-Bromofluorobenzene	12.736	95	161114	44.062	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.120%		
Target Compounds						
16) Acetone	4.296	43	12011	12.020	ug/l #	85
20) Methylene Chloride	5.106	84	6600	2.623	ug/l	97
43) Isopropyl Acetate	8.566	43	15901	2.498	ug/l #	83
95) Naphthalene	15.507	128	6354	5.221	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060521\
Data File : VN067234.D
Acq On : 5 Jun 2021 16:49
Operator : JC/MD
Sample : M2589-37
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FOUNDATION-EXCAVATION-SAMP

Quant Time: Jun 07 01:16:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 04 19:00:20 2021
Response via : Initial Calibration

