

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060521\
 Data File : VN067243.D
 Acq On : 5 Jun 2021 20:20
 Operator : JC/MD
 Sample : M2603-20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FAM4-WC5

Quant Time: Jun 07 01:19:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 04 19:00:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	231095	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.973	114	399617	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	405814	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	177563	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	141441	44.900	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.800%
35) Dibromofluoromethane	8.027	113	116962	47.212	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.420%
50) Toluene-d8	10.446	98	529000	52.431	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.860%
62) 4-Bromofluorobenzene	12.733	95	166139	42.754	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.500%
Target Compounds						
					Qvalue	
16) Acetone	4.309	43	6698	6.162	ug/l	90
20) Methylene Chloride	5.101	84	6520	2.382	ug/l	98
43) Isopropyl Acetate	8.566	43	15444	2.283	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060521\
Data File : VN067243.D
Acq On : 5 Jun 2021 20:20
Operator : JC/MD
Sample : M2603-20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FAM4-WC5

Quant Time: Jun 07 01:19:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 04 19:00:20 2021
Response via : Initial Calibration

