



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060521\
 Data File : VN067221.D
 Acq On : 5 Jun 2021 10:14
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060421W.M
 Title : SW846 8260
 Last Update : Fri Jun 04 19:00:20 2021



AutoFind: Scans 4005, 4006, 4007; Background Corrected with Scan 3990

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	5535	PASS
75	95	30	60	54.5	15421	PASS
95	95	100	100	100.0	28280	PASS
96	95	5	9	6.7	1894	PASS
173	174	0.00	2	0.7	156	PASS
174	95	50	100	78.5	22208	PASS
175	174	5	9	8.0	1782	PASS
176	174	95	101	97.9	21731	PASS
177	176	5	9	7.4	1606	PASS