

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
 Data File : VN078039.D
 Acq On : 05 Jun 2023 13:10
 Operator : JC\MD
 Sample : PB153198TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB153198TB

Quant Time: Jun 06 02:00:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	384511	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	692068	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	643555	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	199119	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	257868	44.821	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.640%
35) Dibromofluoromethane	8.171	113	225747	50.023	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.040%
50) Toluene-d8	10.571	98	877306	50.506	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.020%
62) 4-Bromofluorobenzene	12.853	95	287480	46.860	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.720%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	24209	13.375	ug/l	90
20) Methylene Chloride	5.283	84	14713	1.194	ug/l #	84
37) Ethyl Acetate	7.571	43	64731	11.002	ug/l	97
43) Isopropyl Acetate	8.694	43	288236	29.994	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
Data File : VN078039.D
Acq On : 05 Jun 2023 13:10
Operator : JC\MD
Sample : PB153198TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB153198TB

Quant Time: Jun 06 02:00:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

