

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
 Data File : VN078041.D
 Acq On : 05 Jun 2023 13:57
 Operator : JC\MD
 Sample : PB153198ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB153198ZHE#02

Quant Time: Jun 06 02:00:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	381290	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	713955	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	646964	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	208634	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	257139	45.072	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.140%		
35) Dibromofluoromethane	8.177	113	225595	48.456	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.920%		
50) Toluene-d8	10.571	98	888697	49.593	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.180%		
62) 4-Bromofluorobenzene	12.853	95	262138	41.419	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	82.840%#		
Target Compounds					Qvalue	
16) Acetone	4.442	43	26377	14.696	ug/l	97
20) Methylene Chloride	5.289	84	15300	1.353	ug/l #	89
37) Ethyl Acetate	7.571	43	66222	10.911	ug/l	96
43) Isopropyl Acetate	8.694	43	454330	45.828	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
Data File : VN078041.D
Acq On : 05 Jun 2023 13:57
Operator : JC\MD
Sample : PB153198ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB153198ZHE#02

Quant Time: Jun 06 02:00:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

