

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
 Data File : VN078044.D
 Acq On : 05 Jun 2023 15:09
 Operator : JC\MD
 Sample : PB153198ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB153198ZHE#05

Quant Time: Jun 06 02:02:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	367905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	691209	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	622381	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	194193	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	256691	46.630	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.260%
35) Dibromofluoromethane	8.177	113	222492	49.363	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.720%
50) Toluene-d8	10.571	98	866166	49.926	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.860%
62) 4-Bromofluorobenzene	12.853	95	251272	41.009	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.020%#
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	28324	16.354	ug/l	96
20) Methylene Chloride	5.289	84	15584	1.542	ug/l #	82
37) Ethyl Acetate	7.565	43	69094	11.759	ug/l	97
43) Isopropyl Acetate	8.694	43	467068	48.663	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
Data File : VN078044.D
Acq On : 05 Jun 2023 15:09
Operator : JC\MD
Sample : PB153198ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB153198ZHE#05

Quant Time: Jun 06 02:02:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

