

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
 Data File : VN078045.D
 Acq On : 05 Jun 2023 15:33
 Operator : JC\MD
 Sample : PB153198ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB153198ZHE#06

Quant Time: Jun 06 02:02:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	381752	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	716050	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	673770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	222413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	262491	45.954	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.900%
35) Dibromofluoromethane	8.171	113	230223	49.306	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.620%
50) Toluene-d8	10.571	98	889456	49.490	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.980%
62) 4-Bromofluorobenzene	12.853	95	273633	43.109	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.220%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	29504	16.418	ug/l	99
20) Methylene Chloride	5.289	84	16091	1.525	ug/l #	79
37) Ethyl Acetate	7.571	43	67804	11.139	ug/l #	96
43) Isopropyl Acetate	8.694	43	552274	55.545	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
Data File : VN078045.D
Acq On : 05 Jun 2023 15:33
Operator : JC\MD
Sample : PB153198ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB153198ZHE#06

Quant Time: Jun 06 02:02:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

