

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
 Data File : VN078051.D
 Acq On : 05 Jun 2023 17:57
 Operator : JC\MD
 Sample : PB153198ZHE#12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB153198ZHE#12

Quant Time: Jun 06 02:05:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	377314	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	720278	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	668596	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	223318	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	263911	46.746	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.500%
35) Dibromofluoromethane	8.177	113	230546	49.085	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.180%
50) Toluene-d8	10.576	98	895637	49.541	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.080%
62) 4-Bromofluorobenzene	12.853	95	271858	42.578	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.160%
Target Compounds						
					Qvalue	
16) Acetone	4.459	43	28847	16.241	ug/l	97
37) Ethyl Acetate	7.577	43	66515	10.863	ug/l	98
43) Isopropyl Acetate	8.694	43	596937	59.684	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
Data File : VN078051.D
Acq On : 05 Jun 2023 17:57
Operator : JC\MD
Sample : PB153198ZHE#12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB153198ZHE#12

Quant Time: Jun 06 02:05:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

