

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
 Data File : VN078055.D
 Acq On : 05 Jun 2023 19:33
 Operator : JC\MD
 Sample : 03018-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-11991

Quant Time: Jun 06 02:06:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	383340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	709444	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	669723	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	215735	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	263932	46.015	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.040%
35) Dibromofluoromethane	8.177	113	232047	50.159	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.320%
50) Toluene-d8	10.571	98	906279	50.896	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.800%
62) 4-Bromofluorobenzene	12.853	95	301739	47.979	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.960%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	20906	11.585	ug/l #	82
20) Methylene Chloride	5.294	84	14003	1.047	ug/l	97
37) Ethyl Acetate	7.571	43	64025	10.616	ug/l #	93
43) Isopropyl Acetate	8.694	43	321400	32.626	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
Data File : VN078055.D
Acq On : 05 Jun 2023 19:33
Operator : JC\MD
Sample : 03018-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-11991

Quant Time: Jun 06 02:06:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

