

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
 Data File : VN078056.D
 Acq On : 05 Jun 2023 19:57
 Operator : JC\MD
 Sample : 03018-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 278

Quant Time: Jun 06 02:07:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	372947	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	698768	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	633761	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	197436	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	258612	46.344	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.680%
35) Dibromofluoromethane	8.171	113	223425	49.033	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.060%
50) Toluene-d8	10.571	98	870419	49.629	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.260%
62) 4-Bromofluorobenzene	12.853	95	280705	45.317	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.640%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	22777	12.974	ug/l	96
20) Methylene Chloride	5.294	84	17419	1.912	ug/l	91
37) Ethyl Acetate	7.571	43	50651	8.527	ug/l	94
43) Isopropyl Acetate	8.694	43	281291	28.990	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
Data File : VN078056.D
Acq On : 05 Jun 2023 19:57
Operator : JC\MD
Sample : 03018-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
278

Quant Time: Jun 06 02:07:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

