

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
 Data File : VN078057.D
 Acq On : 05 Jun 2023 20:21
 Operator : JC\MD
 Sample : 03018-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RT-4648

Quant Time: Jun 06 02:07:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	372879	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	697770	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	635412	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	197907	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	257305	46.118	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.240%
35) Dibromofluoromethane	8.171	113	227075	49.906	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.820%
50) Toluene-d8	10.577	98	881837	50.352	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.700%
62) 4-Bromofluorobenzene	12.853	95	276422	44.689	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.380%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	20965	11.944	ug/l	94
20) Methylene Chloride	5.289	84	15501	1.475	ug/l #	82
37) Ethyl Acetate	7.571	43	50287	8.477	ug/l	97
43) Isopropyl Acetate	8.694	43	279382	28.835	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
Data File : VN078057.D
Acq On : 05 Jun 2023 20:21
Operator : JC\MD
Sample : 03018-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RT-4648

Quant Time: Jun 06 02:07:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

