

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
 Data File : VN078058.D
 Acq On : 05 Jun 2023 20:44
 Operator : JC\MD
 Sample : 03029-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-6

Quant Time: Jun 06 02:08:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	371827	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	705359	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	652579	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	208112	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	260651	46.850	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.700%
35) Dibromofluoromethane	8.177	113	226464	49.236	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.480%
50) Toluene-d8	10.571	98	883336	49.895	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.780%
62) 4-Bromofluorobenzene	12.853	95	289554	46.309	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.620%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	23906	13.658	ug/l #	82
20) Methylene Chloride	5.294	84	16391	1.689	ug/l #	87
37) Ethyl Acetate	7.577	43	62435	10.412	ug/l	97
43) Isopropyl Acetate	8.694	43	397489	40.583	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
Data File : VN078058.D
Acq On : 05 Jun 2023 20:44
Operator : JC\MD
Sample : 03029-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-6

Quant Time: Jun 06 02:08:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

