

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
 Data File : VN078059.D
 Acq On : 05 Jun 2023 21:08
 Operator : JC\MD
 Sample : 03029-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-7

Quant Time: Jun 06 02:08:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	379282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	708859	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	664854	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	213620	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	257322	45.343	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.680%
35) Dibromofluoromethane	8.177	113	223095	48.264	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.520%
50) Toluene-d8	10.571	98	902053	50.700	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.400%
62) 4-Bromofluorobenzene	12.853	95	291541	46.396	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.800%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	25901	14.507	ug/l	91
20) Methylene Chloride	5.283	84	15731	1.467	ug/l #	95
37) Ethyl Acetate	7.577	43	66537	11.041	ug/l	97
43) Isopropyl Acetate	8.700	43	439983	44.700	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060523\
Data File : VN078059.D
Acq On : 05 Jun 2023 21:08
Operator : JC\MD
Sample : 03029-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-7

Quant Time: Jun 06 02:08:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

