

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060524\
 Data File : VN082460.D
 Acq On : 05 Jun 2024 10:49
 Operator : JC\MD
 Sample : VN0605MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0605MBL01

Quant Time: Jun 06 00:52:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	147506	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	267348	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	257681	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	104593	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	127656	54.808	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.620%
35) Dibromofluoromethane	8.171	113	91706	55.762	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.520%
50) Toluene-d8	10.571	98	342978	54.395	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.800%
62) 4-Bromofluorobenzene	12.853	95	110847	49.254	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	98.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060524\
Data File : VN082460.D
Acq On : 05 Jun 2024 10:49
Operator : JC\MD
Sample : VN0605MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0605MBL01

Quant Time: Jun 06 00:52:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

