

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060524\
 Data File : VN082461.D
 Acq On : 05 Jun 2024 11:14
 Operator : JC\MD
 Sample : VN0605WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0605WBL01

Quant Time: Jun 06 00:52:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	157039	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	288593	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	279042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	118887	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	137972	55.641	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.280%
35) Dibromofluoromethane	8.171	113	99992	56.324	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.640%
50) Toluene-d8	10.571	98	372625	54.747	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.500%
62) 4-Bromofluorobenzene	12.847	95	122797	50.547	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	101.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060524\
Data File : VN082461.D
Acq On : 05 Jun 2024 11:14
Operator : JC\MD
Sample : VN0605WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0605WBL01

Quant Time: Jun 06 00:52:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

