

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060524\
 Data File : VN082472.D
 Acq On : 05 Jun 2024 15:49
 Operator : JC\MD
 Sample : P2704-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WCS-TPN

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/06/2024
 Supervised By : Mahesh Dadoda 06/06/2024

Quant Time: Jun 06 00:58:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	148142	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	282328	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	282550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115368	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	134686	57.578	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 115.160%			
35) Dibromofluoromethane	8.171	113	96401	55.506	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 111.020%			
50) Toluene-d8	10.571	98	376804	56.589	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 113.180%#			
62) 4-Bromofluorobenzene	12.853	95	125711	52.895	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 105.800%			
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	103529	107.890	ug/l	92
20) Methylene Chloride	5.277	84	3757	1.941	ug/l #	66
25) 2-Butanone	7.489	43	24850	17.080	ug/l #	80
43) Isopropyl Acetate	8.694	43	263265	45.578	ug/l #	72
51) 4-Methyl-2-Pentanone	10.447	43	22414m	6.830	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060524\
Data File : VN082472.D
Acq On : 05 Jun 2024 15:49
Operator : JC\MD
Sample : P2704-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WCS-TPN

Quant Time: Jun 06 00:58:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/06/2024
Supervised By : Mahesh Dadoda 06/06/2024

