

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060524\
 Data File : VN082475.D
 Acq On : 05 Jun 2024 17:02
 Operator : JC\MD
 Sample : P2704-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WCS-TPO

Quant Time: Jun 06 00:59:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/06/2024
 Supervised By : Mahesh Dadoda 06/06/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	159619	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	295985	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	296249	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	112691	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	129092	51.218	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.440%			
35) Dibromofluoromethane	8.171	113	94397	51.845	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.680%			
50) Toluene-d8	10.570	98	365480	52.356	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 104.720%			
62) 4-Bromofluorobenzene	12.853	95	116985	46.952	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 93.900%			
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	100751	97.445	ug/l	91
20) Methylene Chloride	5.277	84	4390	2.104	ug/l #	82
25) 2-Butanone	7.488	43	25210	16.082	ug/l	92
43) Isopropyl Acetate	8.694	43	268521	44.343	ug/l #	73
51) 4-Methyl-2-Pentanone	10.447	43	23843m	6.930	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060524\
Data File : VN082475.D
Acq On : 05 Jun 2024 17:02
Operator : JC\MD
Sample : P2704-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WCS-TPO

Quant Time: Jun 06 00:59:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/06/2024
Supervised By : Mahesh Dadoda 06/06/2024

