

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060524\
 Data File : VN082461.D
 Acq On : 05 Jun 2024 11:14
 Operator : JC\MD
 Sample : VN0605WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0605WBL01

Quant Time: Jun 06 00:52:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	157039	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	288593	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	279042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	118887	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	137972	55.641	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.280%
35) Dibromofluoromethane	8.171	113	99992	56.324	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.640%
50) Toluene-d8	10.571	98	372625	54.747	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.500%
62) 4-Bromofluorobenzene	12.847	95	122797	50.547	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	101.100%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060524\
Data File : VN082461.D
Acq On : 05 Jun 2024 11:14
Operator : JC\MD
Sample : VN0605WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0605WBL01

Quant Time: Jun 06 00:52:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN082461.D
Acq: 05 Jun 2024 11:14




Instrument :
MSVOA_N
ClientSampleId :
VN0605WBL01


Tgt Ion:168 Resp: 157039
Ion Ratio Lower Upper
168 100
99 58.4 48.2 72.4



#33
1,2-Dichloroethane-d4
Concen: 55.641 ug/l
RT: 8.583 min Scan# 1127
Delta R.T. 0.006 min
Lab File: VN082461.D
Acq: 05 Jun 2024 11:14

Tgt Ion: 65 Resp: 137972
Ion Ratio Lower Upper
65 100
67 52.0 0.0 103.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN082461.D

Acq: 05 Jun 2024 11:14

Instrument :

MSVOA_N

ClientSampleId :

VN0605WBL01





Tgt Ion:114 Resp: 288593

Ion Ratio Lower Upper

114 100

63 23.3 0.0 41.0

88 16.2 0.0 31.4



#35

Dibromofluoromethane

Concen: 56.324 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN082461.D

Acq: 05 Jun 2024 11:14




Tgt Ion:113 Resp: 99992


Ion Ratio Lower Upper

113 100

111 104.5 83.4 125.2

192 18.0 14.0 21.0









#50
Toluene-d8
Concen: 54.747 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. 0.000 min
Lab File: VN082461.D
Acq: 05 Jun 2024 11:14




Instrument :
MSVOA_N
ClientSampleId :
VN0605WBL01


Tgt Ion: 98 Resp: 372625
Ion Ratio Lower Upper
98 100
100 64.2 51.8 77.6



#62
4-Bromofluorobenzene
Concen: 50.547 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN082461.D
Acq: 05 Jun 2024 11:14

Tgt Ion: 95 Resp: 122797
Ion Ratio Lower Upper
95 100
174 78.6 0.0 160.6
176 75.2 0.0 152.8








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN082461.D
Acq: 05 Jun 2024 11:14

Instrument :
MSVOA_N
ClientSampleId :
VN0605WBL01

Tgt Ion:117 Resp: 279042
Ion Ratio Lower Upper
117 100
82 60.1 44.6 66.8
119 30.9 25.9 38.9



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN082461.D
Acq: 05 Jun 2024 11:14

Tgt Ion:152 Resp: 118887
Ion Ratio Lower Upper
152 100
115 59.2 30.4 91.3
150 157.8 0.0 353.2

