

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060619\
 Data File : VN056024.D
 Acq On : 5 Jun 2019 23:28
 Operator : JC/SP
 Sample : K3182-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-10

Quant Time: Jun 06 04:38:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N060619W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jun 06 04:02:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	54695	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	293818	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	238110	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	102931	30.30	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.00%
60) 4-Bromofluorobenzene	12.40	95	83793	23.74	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	79.13%
63) Toluene-d8	10.09	98	340900	31.56	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	105.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060619\
Data File : VN056024.D
Acq On : 5 Jun 2019 23:28
Operator : JC/SP
Sample : K3182-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-10

Quant Time: Jun 06 04:38:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N060619W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jun 06 04:02:53 2019
Response via : Initial Calibration

