

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060622\
 Data File : VN072671.D
 Acq On : 06 Jun 2022 18:41
 Operator : JC\MD
 Sample : N3123-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-2

Quant Time: Jun 07 06:29:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N060622W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jun 06 14:10:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.651	128	40135	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.963	114	238143	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.739	117	193239	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.439	65	140364	30.665	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.233%
60) 4-Bromofluorobenzene	12.733	95	97362	24.497	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	81.667%
63) Toluene-d8	10.439	98	350528	32.746	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	109.167%
Target Compounds						
					Qvalue	
25) Chloroform	7.816	83	65658	10.141	ug/l	96
41) Bromodichloromethane	9.763	83	5950	1.098	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060622\
Data File : VN072671.D
Acq On : 06 Jun 2022 18:41
Operator : JC\MD
Sample : N3123-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
14B-2

Quant Time: Jun 07 06:29:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N060622W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jun 06 14:10:15 2022
Response via : Initial Calibration

