

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060622\
 Data File : VN072672.D
 Acq On : 06 Jun 2022 19:06
 Operator : JC\MD
 Sample : N3123-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-3

Quant Time: Jun 07 06:29:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N060622W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jun 06 14:10:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.657	128	40645	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.963	114	235064	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.745	117	198857	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.434	65	138030	29.777	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.267%
60) 4-Bromofluorobenzene	12.727	95	102253	25.001	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	83.333%
63) Toluene-d8	10.439	98	350643	31.831	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	106.100%
Target Compounds						
					Qvalue	
15) Acetone	4.293	58	4101	6.707	ug/l	83
25) Chloroform	7.816	83	62160	9.480	ug/l	91
41) Bromodichloromethane	9.763	83	5545	1.037	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060622\
Data File : VN072672.D
Acq On : 06 Jun 2022 19:06
Operator : JC\MD
Sample : N3123-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
14B-3

Quant Time: Jun 07 06:29:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N060622W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jun 06 14:10:15 2022
Response via : Initial Calibration

