

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060623\
 Data File : VN078065.D
 Acq On : 06 Jun 2023 12:39
 Operator : JC\MD
 Sample : VN0606WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0606WBL01

Quant Time: Jun 07 01:20:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	384151	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	702804	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	606986	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	178448	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	260956	45.400	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.800%
35) Dibromofluoromethane	8.171	113	223175	48.697	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.400%
50) Toluene-d8	10.571	98	864247	48.994	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.980%
62) 4-Bromofluorobenzene	12.853	95	267057	42.866	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060623\
Data File : VN078065.D
Acq On : 06 Jun 2023 12:39
Operator : JC\MD
Sample : VN0606WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0606WBL01

Quant Time: Jun 07 01:20:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

