

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060623\
 Data File : VN078072.D
 Acq On : 06 Jun 2023 16:00
 Operator : JC\MD
 Sample : 03029-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-9

Quant Time: Jun 07 01:22:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	427228	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	794241	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	711071	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	230592	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	276593	43.269	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.540%
35) Dibromofluoromethane	8.177	113	247438	47.776	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.560%
50) Toluene-d8	10.577	98	973060	48.812	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.620%
62) 4-Bromofluorobenzene	12.853	95	330343	46.920	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.840%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	29864	14.849	ug/l	93
20) Methylene Chloride	5.289	84	19096	1.741	ug/l	90
37) Ethyl Acetate	7.571	43	78665	11.651	ug/l	98
43) Isopropyl Acetate	8.700	43	335169	30.391	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060623\
Data File : VN078072.D
Acq On : 06 Jun 2023 16:00
Operator : JC\MD
Sample : 03029-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-9

Quant Time: Jun 07 01:22:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

