

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060623\
 Data File : VN078073.D
 Acq On : 06 Jun 2023 16:24
 Operator : JC\MD
 Sample : 03029-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-10

Quant Time: Jun 07 01:22:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	445080	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	813468	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	742707	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	243497	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	290267	43.587	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.180%
35) Dibromofluoromethane	8.171	113	251945	47.496	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.000%
50) Toluene-d8	10.571	98	1008114	49.375	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.740%
62) 4-Bromofluorobenzene	12.853	95	336513	46.666	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.340%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	28323	13.518	ug/l	97
20) Methylene Chloride	5.289	84	22274	2.195	ug/l	90
37) Ethyl Acetate	7.577	43	76966	11.130	ug/l	97
43) Isopropyl Acetate	8.694	43	326362	28.893	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060623\
Data File : VN078073.D
Acq On : 06 Jun 2023 16:24
Operator : JC\MD
Sample : 03029-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-10

Quant Time: Jun 07 01:22:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

