

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060623\
 Data File : VN078076.D
 Acq On : 06 Jun 2023 17:36
 Operator : JC\MD
 Sample : 03036-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-STONE-0602

Quant Time: Jun 07 01:23:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	422104	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	771847	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	700282	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	218478	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	278823	44.147	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.300%
35) Dibromofluoromethane	8.177	113	241836	48.049	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.100%
50) Toluene-d8	10.576	98	952016	49.142	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.280%
62) 4-Bromofluorobenzene	12.853	95	320456	46.836	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.680%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	26743	13.459	ug/l #	77
20) Methylene Chloride	5.294	84	22009	2.374	ug/l #	91
37) Ethyl Acetate	7.571	43	63166	9.627	ug/l	97
43) Isopropyl Acetate	8.694	43	342582	31.964	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060623\
Data File : VN078076.D
Acq On : 06 Jun 2023 17:36
Operator : JC\MD
Sample : 03036-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-STONE-0602

Quant Time: Jun 07 01:23:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

