

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060623\
 Data File : VN078085.D
 Acq On : 06 Jun 2023 23:46
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 07 05:05:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	396209	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	720115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	609976	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	175980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	254474	48.171	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.340%
35) Dibromofluoromethane	8.177	113	224237	49.248	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.500%
50) Toluene-d8	10.571	98	857472	48.850	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.700%
62) 4-Bromofluorobenzene	12.853	95	238623	43.729	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.460%
Target Compounds						
3) Chloromethane	2.377	50	1361	Below Cal	Qvalue	99
75) 1,1,2,2-Tetrachloroethane	12.947	83	160	Below Cal	#	69
94) Hexachlorobutadiene	15.506	225	382	Below Cal		82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060623\
Data File : VN078085.D
Acq On : 06 Jun 2023 23:46
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 07 05:05:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

