

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060623\
 Data File : VN078063.D
 Acq On : 06 Jun 2023 11:39
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 07 01:19:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.569	0.455	20.0	65	0.00
3 P	Chloromethane	0.682	0.516	24.3	72	0.00
4 C	Vinyl Chloride	0.717	0.665	7.3#	78	0.01
5 T	Bromomethane	0.531	0.514	3.2	87	0.00
6 T	Chloroethane	0.517	0.511	1.2	88	0.00
7 T	Trichlorofluoromethane	1.000	1.000	0.0	86	0.00
8 T	Diethyl Ether	0.365	0.373	-2.2	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.553	0.570	-3.1	90	0.00
10 T	Methyl Iodide	0.660	0.781	-18.3	101	0.00
11 T	Tert butyl alcohol	0.103	0.097	5.8	82	0.00
12 CM	1,1-Dichloroethene	0.526	0.528	-0.4#	86	0.00
13 T	Acrolein	0.073	0.062	15.1	78	0.00
14 T	Allyl chloride	0.773	0.744	3.8	89	0.00
15 T	Acrylonitrile	0.276	0.277	-0.4	88	0.01
16 T	Acetone	0.235	0.262	-11.5	102	0.00
17 T	Carbon Disulfide	1.417	1.138	19.7	73	0.01
18 T	Methyl Acetate	0.987	0.981	0.6	87	0.00
19 T	Methyl tert-butyl Ether	1.904	2.015	-5.8	90	0.00
20 T	Methylene Chloride	0.705	0.662	6.1	92	0.00
21 T	trans-1,2-Dichloroethene	0.591	0.604	-2.2	88	0.00
22 T	Diisopropyl ether	1.811	1.897	-4.7	91	0.00
23 T	Vinyl Acetate	1.122	1.121	0.1	86	0.00
24 P	1,1-Dichloroethane	1.109	1.195	-7.8	93	0.00
25 T	2-Butanone	0.368	0.371	-0.8	90	0.00
26 T	2,2-Dichloropropane	0.943	1.056	-12.0	91	0.00
27 T	cis-1,2-Dichloroethene	0.708	0.772	-9.0	94	0.00
28 T	Bromochloromethane	0.480	0.458	4.6	82	0.00
29 T	Tetrahydrofuran	0.241	0.225	6.6	83	0.00
30 C	Chloroform	1.220	1.321	-8.3#	96	0.00
31 T	Cyclohexane	1.017	0.864	15.0	77	0.00
32 T	1,1,1-Trichloroethane	1.038	1.145	-10.3	93	0.00
33 S	1,2-Dichloroethane-d4	0.748	0.758	-1.3	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.326	0.384	-17.8	95	0.00
36 T	1,1-Dichloropropene	0.477	0.530	-11.1	90	0.00
37 T	Ethyl Acetate	0.425	0.406	4.5	83	0.00
38 T	Carbon Tetrachloride	0.482	0.570	-18.3	95	0.00
39 T	Methylcyclohexane	0.584	0.520	11.0	72	0.00
40 TM	Benzene	1.407	1.564	-11.2	92	0.00
41 T	Methacrylonitrile	0.226	0.223	1.3	84	0.00
42 TM	1,2-Dichloroethane	0.511	0.548	-7.2	91	0.00
43 T	Isopropyl Acetate	0.694	0.744	-7.2	90	0.00
44 TM	Trichloroethene	0.356	0.410	-15.2	96	0.00
45 C	1,2-Dichloropropane	0.359	0.410	-14.2#	93	0.00
46 T	Dibromomethane	0.257	0.283	-10.1	93	0.00
47 T	Bromodichloromethane	0.492	0.585	-18.9	96	0.00
48 T	Methyl methacrylate	0.329	0.343	-4.3	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060623\
 Data File : VN078063.D
 Acq On : 06 Jun 2023 11:39
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 07 01:19:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.007	-16.7	88	0.00
50 S	Toluene-d8	1.255	1.364	-8.7	87	0.00
51 T	4-Methyl-2-Pentanone	0.413	0.435	-5.3	86	0.00
52 CM	Toluene	0.876	1.024	-16.9#	94	0.00
53 T	t-1,3-Dichloropropene	0.521	0.616	-18.2	94	0.00
54 T	cis-1,3-Dichloropropene	0.574	0.674	-17.4	93	0.00
55 T	1,1,2-Trichloroethane	0.348	0.411	-18.1	99	0.00
56 T	Ethyl methacrylate	0.526	0.590	-12.2	90	0.00
57 T	1,3-Dichloropropane	0.596	0.675	-13.3	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.198	23.6	62	0.00
59 T	2-Hexanone	0.305	0.314	-3.0	84	0.00
60 T	Dibromochloromethane	0.359	0.448	-24.8	99	0.00
61 T	1,2-Dibromoethane	0.347	0.405	-16.7	92	0.00
62 S	4-Bromofluorobenzene	0.443	0.467	-5.4	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	80	0.00
64 T	Tetrachloroethene	0.350	0.372	-6.3	87	0.00
65 PM	Chlorobenzene	1.084	1.283	-18.4	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.375	0.491	-30.9#	100	0.00
67 C	Ethyl Benzene	1.942	2.280	-17.4#	92	0.00
68 T	m/p-Xylenes	0.736	0.878	-19.3	94	0.00
69 T	o-Xylene	0.724	0.861	-18.9	93	0.00
70 T	Styrene	1.150	1.420	-23.5	95	0.00
71 P	Bromoform	0.250	0.323	-29.2#	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	4.310	5.063	-17.5	91	0.00
74 T	N-amyl acetate	1.593	1.557	2.3	79	0.00
75 P	1,1,2,2-Tetrachloroethane	1.326	1.504	-13.4	93	0.00
76 T	1,2,3-Trichloropropane	1.252	1.185	5.4	77	0.00
77 T	Bromobenzene	0.984	1.141	-16.0	95	0.00
78 T	n-propylbenzene	5.081	5.702	-12.2	89	0.00
79 T	2-Chlorotoluene	3.032	3.451	-13.8	91	0.00
80 T	1,3,5-Trimethylbenzene	3.622	4.171	-15.2	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.392	0.432	-10.2	87	0.00
82 T	4-Chlorotoluene	3.017	3.340	-10.7	91	0.00
83 T	tert-Butylbenzene	3.162	3.567	-12.8	88	0.00
84 T	1,2,4-Trimethylbenzene	3.563	4.187	-17.5	93	0.00
85 T	sec-Butylbenzene	4.494	4.714	-4.9	81	0.00
86 T	p-Isopropyltoluene	3.583	3.913	-9.2	84	0.00
87 T	1,3-Dichlorobenzene	1.819	2.001	-10.0	90	0.00
88 T	1,4-Dichlorobenzene	1.805	1.970	-9.1	89	0.00
89 T	n-Butylbenzene	3.296	3.172	3.8	75	0.00
90 T	Hexachloroethane	0.589	0.705	-19.7	88	0.00
91 T	1,2-Dichlorobenzene	1.699	2.004	-18.0	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.250	2.7	79	0.00
93 T	1,2,4-Trichlorobenzene	0.993	0.785	20.9	64	0.00
94 T	Hexachlorobutadiene	0.412	0.381	7.5	71	0.00
95 T	Naphthalene	3.418	2.442	28.6#	59	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060623\
Data File : VN078063.D
Acq On : 06 Jun 2023 11:39
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 07 01:19:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.972	0.808	16.9	66	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6