

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086862.D
 Acq On : 06 Jun 2025 12:44
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/09/2025
 Supervised By : Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:54:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	212682	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	393457	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	348010	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	167399	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	0.000%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.153	85	1985	0.936	ug/l	72
3) Chloromethane	2.400	50	3241	1.184	ug/l	98
4) Vinyl Chloride	2.559	62	2848	1.008	ug/l #	88
6) Chloroethane	3.159	64	1957	1.073	ug/l #	79
7) Trichlorofluoromethane	3.530	101	3752	1.017	ug/l	97
8) Diethyl Ether	3.994	74	1583	0.984	ug/l	76
9) 1,1,2-Trichlorotrifluo...	4.377	101	2358m	1.017	ug/l	
12) 1,1-Dichloroethene	4.365	96	2436	1.029	ug/l	85
14) Allyl chloride	5.041	41	4189	1.067	ug/l #	75
15) Acrylonitrile	5.735	53	9222	5.106	ug/l	96
16) Acetone	4.459	43	9069	6.006	ug/l	95
17) Carbon Disulfide	4.730	76	7306	1.116	ug/l	97
18) Methyl Acetate	5.053	43	4402	1.000	ug/l #	66
19) Methyl tert-butyl Ether	5.824	73	9019	1.052	ug/l	92
20) Methylene Chloride	5.294	84	3496	1.237	ug/l	93
21) trans-1,2-Dichloroethene	5.800	96	2979	1.131	ug/l	87
22) Diisopropyl ether	6.682	45	8584	1.037	ug/l #	96
23) Vinyl Acetate	6.618	43	37075	5.303	ug/l	95
24) 1,1-Dichloroethane	6.588	63	5071	1.065	ug/l #	81
25) 2-Butanone	7.500	43	12841	5.231	ug/l	93
26) 2,2-Dichloropropane	7.500	77	4112	1.110	ug/l	94
27) cis-1,2-Dichloroethene	7.494	96	3345	1.062	ug/l	95
28) Bromochloromethane	7.824	49	2464	1.052	ug/l #	88
29) Tetrahydrofuran	7.853	42	8301	5.191	ug/l	100
30) Chloroform	7.971	83	5255	1.105	ug/l	97
32) 1,1,1-Trichloroethane	8.171	97	4378	1.082	ug/l #	50
36) 1,1-Dichloropropene	8.376	75	3678	1.059	ug/l	96
37) Ethyl Acetate	7.577	43	4282	0.968	ug/l #	78
38) Carbon Tetrachloride	8.376	117	3563	1.045	ug/l #	93
39) Methylcyclohexane	9.606	83	4981	1.046	ug/l #	90
40) Benzene	8.612	78	12498	1.100	ug/l	100
41) Methacrylonitrile	7.782	41	2799	1.127	ug/l	89
42) 1,2-Dichloroethane	8.682	62	3725	1.081	ug/l	82
43) Isopropyl Acetate	8.694	43	7673	1.081	ug/l #	96
44) Trichloroethene	9.365	130	2822	1.047	ug/l	71
45) 1,2-Dichloropropane	9.629	63	2883	1.043	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086862.D
 Acq On : 06 Jun 2025 12:44
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/09/2025
 Supervised By : Mahesh Dadoda 06/09/2025

Quant Time: Jun 07 01:54:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 01:51:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	9.712	93	1835	0.999	ug/l	94
47) Bromodichloromethane	9.888	83	4017	1.063	ug/l #	96
48) Methyl methacrylate	9.688	41	3492	1.069	ug/l	88
49) 1,4-Dioxane	9.706	88	1004	16.891	ug/l #	74
51) 4-Methyl-2-Pentanone	10.453	43	19886	4.653	ug/l	99
52) Toluene	10.635	92	7227	1.041	ug/l	95
53) t-1,3-Dichloropropene	10.841	75	4494	1.064	ug/l	97
54) cis-1,3-Dichloropropene	10.312	75	4792	1.060	ug/l	97
55) 1,1,2-Trichloroethane	11.023	97	2828	1.058	ug/l	89
56) Ethyl methacrylate	10.882	69	3406	0.800	ug/l #	85
57) 1,3-Dichloropropane	11.170	76	5062	1.092	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.165	63	10924	4.304	ug/l	97
59) 2-Hexanone	11.235	43	12295m	4.466	ug/l	
60) Dibromochloromethane	11.359	129	2843	1.021	ug/l	96
61) 1,2-Dibromoethane	11.470	107	2777	1.014	ug/l	96
64) Tetrachloroethene	11.106	164	2473	1.123	ug/l	94
65) Chlorobenzene	11.894	112	8581	1.118	ug/l #	87
66) 1,1,1,2-Tetrachloroethane	11.959	131	2523	1.023	ug/l #	63
67) Ethyl Benzene	11.970	91	13845	1.047	ug/l	94
68) m/p-Xylenes	12.070	106	10163	2.008	ug/l	95
69) o-Xylene	12.400	106	4970	1.025	ug/l	93
70) Styrene	12.417	104	8189	0.987	ug/l	98
71) Bromoform	12.582	173	1508	0.825	ug/l #	78
73) Isopropylbenzene	12.694	105	12938	1.061	ug/l	98
74) N-amyl acetate	12.647	43	4608m	1.138	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.935	83	4326	1.047	ug/l #	93
76) 1,2,3-Trichloropropane	13.000	75	4343m	1.088	ug/l	
77) Bromobenzene	12.982	156	2913	1.042	ug/l	96
78) n-propylbenzene	13.035	91	15168	1.023	ug/l	98
79) 2-Chlorotoluene	13.123	91	9550	1.074	ug/l	97
80) 1,3,5-Trimethylbenzene	13.170	105	10058	0.999	ug/l	96
82) 4-Chlorotoluene	13.223	91	9706	1.079	ug/l	96
83) tert-Butylbenzene	13.435	119	9849	1.069	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	9882	0.979	ug/l	99
85) sec-Butylbenzene	13.617	105	13775	1.029	ug/l	100
86) p-Isopropyltoluene	13.729	119	11065	1.000	ug/l	97
87) 1,3-Dichlorobenzene	13.735	146	5903	1.073	ug/l	94
88) 1,4-Dichlorobenzene	13.811	146	6094m	1.086	ug/l	
89) n-Butylbenzene	14.058	91	11627	1.085	ug/l	94
90) Hexachloroethane	14.329	117	1942	1.036	ug/l	93
91) 1,2-Dichlorobenzene	14.106	146	5607	1.061	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	1136	1.150	ug/l	78
93) 1,2,4-Trichlorobenzene	15.388	180	3489	1.034	ug/l	98
94) Hexachlorobutadiene	15.500	225	1219	0.969	ug/l	88
95) Naphthalene	15.635	128	12789	1.018	ug/l	98
96) 1,2,3-Trichlorobenzene	15.835	180	3594	1.072	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

