

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086871.D
 Acq On : 06 Jun 2025 17:10
 Operator : JC\MD
 Sample : VN0606WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0606WBL02

Quant Time: Jun 07 02:25:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	344048	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	613851	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	529685	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	242229	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	217725	47.266	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.540%
35) Dibromofluoromethane	8.177	113	177475	48.786	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.580%
50) Toluene-d8	10.571	98	740500	51.419	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.840%
62) 4-Bromofluorobenzene	12.847	95	254971	47.654	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086871.D
Acq On : 06 Jun 2025 17:10
Operator : JC\MD
Sample : VN0606WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0606WBL02

Quant Time: Jun 07 02:25:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

