

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086879.D
 Acq On : 06 Jun 2025 20:13
 Operator : JC\MD
 Sample : Q2216-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 3865

Quant Time: Jun 07 02:30:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	334776	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	602082	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	526101	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	256470	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	208275	46.466	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.940%
35) Dibromofluoromethane	8.177	113	172503	48.346	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.700%
50) Toluene-d8	10.571	98	729350	51.635	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.280%
62) 4-Bromofluorobenzene	12.847	95	263968	50.300	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.600%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.547	59	25417	20.898	ug/l	95
16) Acetone	4.447	43	9389	3.950	ug/l	97
95) Naphthalene	15.635	128	44948	2.335	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086879.D
Acq On : 06 Jun 2025 20:13
Operator : JC\MD
Sample : Q2216-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
3865

Quant Time: Jun 07 02:30:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

