

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086884.D
 Acq On : 06 Jun 2025 22:07
 Operator : JC\MD
 Sample : Q2198-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B-202-SB02

Quant Time: Jun 07 02:32:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	409861	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	701621	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	607667	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	296022	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	229171	41.762	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.520%
35) Dibromofluoromethane	8.177	113	199395	47.955	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.900%
50) Toluene-d8	10.571	98	832222	50.559	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.120%
62) 4-Bromofluorobenzene	12.847	95	298767	48.854	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.700%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	74188	25.493	ug/l	100
20) Methylene Chloride	5.300	84	18722	3.436	ug/l	96
43) Isopropyl Acetate	8.694	43	27252	2.153	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086884.D
Acq On : 06 Jun 2025 22:07
Operator : JC\MD
Sample : Q2198-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B-202-SB02

Quant Time: Jun 07 02:32:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

