

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086885.D
 Acq On : 06 Jun 2025 22:29
 Operator : JC\MD
 Sample : Q2198-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B-207-SB02

Quant Time: Jun 07 02:33:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	403626	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	688747	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	589655	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	289711	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	222690	41.208	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	82.420%		
35) Dibromofluoromethane	8.177	113	195297	47.847	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.700%		
50) Toluene-d8	10.571	98	818506	50.656	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	101.320%		
62) 4-Bromofluorobenzene	12.847	95	291833	48.612	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	97.220%		
Target Compounds					Qvalue	
16) Acetone	4.453	43	84930	29.635	ug/l	96
20) Methylene Chloride	5.300	84	17613	3.283	ug/l	92
43) Isopropyl Acetate	8.700	43	22190	1.786	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086885.D
Acq On : 06 Jun 2025 22:29
Operator : JC\MD
Sample : Q2198-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B-207-SB02

Quant Time: Jun 07 02:33:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

