

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060719\
 Data File : VN056041.D
 Acq On : 6 Jun 2019 13:15
 Operator : JC/SP
 Sample : K3195-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LF014RW040-190603

Quant Time: Jun 06 19:07:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	336261	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	506268	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	466720	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	148359	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	179024	53.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.24%	
35) Dibromofluoromethane	7.59	113	156409	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
50) Toluene-d8	10.09	98	643565	53.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.68%	
62) 4-Bromofluorobenzene	12.40	95	199842	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.83	96	28245	7.944	ug/l	87
44) Trichloroethene	8.83	130	16463	4.333	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060719\
Data File : VN056041.D
Acq On : 6 Jun 2019 13:15
Operator : JC/SP
Sample : K3195-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LF014RW040-190603

Quant Time: Jun 06 19:07:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

