

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060719\
 Data File : VN056042.D
 Acq On : 6 Jun 2019 13:40
 Operator : JC/SP
 Sample : K3195-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LF014MW009-190603

Quant Time: Jun 06 19:11:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	340215	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	514352	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	475973	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	147221	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	176813	52.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.68%	
35) Dibromofluoromethane	7.59	113	154857	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
50) Toluene-d8	10.09	98	643844	53.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.04%	
62) 4-Bromofluorobenzene	12.40	95	199681	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
Target Compounds						
4) Vinyl Chloride	2.18	62	23250	6.285	ug/l	99
27) cis-1,2-Dichloroethene	6.83	96	34377	9.557	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060719\
Data File : VN056042.D
Acq On : 6 Jun 2019 13:40
Operator : JC/SP
Sample : K3195-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LF014MW009-190603

Quant Time: Jun 06 19:11:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

