

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060719\
 Data File : VN056046.D
 Acq On : 6 Jun 2019 15:20
 Operator : JC/SP
 Sample : K3195-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP-0529-190603

Quant Time: Jun 07 05:52:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	333138	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	503046	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	468054	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	145679	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	177707	53.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.44%	
35) Dibromofluoromethane	7.59	113	155756	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
50) Toluene-d8	10.09	98	647236	54.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.00%	
62) 4-Bromofluorobenzene	12.40	95	198855	49.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.48%	
Target Compounds						
4) Vinyl Chloride	2.19	62	21113	5.828	ug/l	97
27) cis-1,2-Dichloroethene	6.83	96	32536	9.237	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060719\
Data File : VN056046.D
Acq On : 6 Jun 2019 15:20
Operator : JC/SP
Sample : K3195-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP-0529-190603

Quant Time: Jun 07 05:52:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

