

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN060719\
 Data File : VN056051.D
 Acq On : 6 Jun 2019 17:26
 Operator : JC/SP
 Sample : K3203-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LF015MW100-190604

Quant Time: Jun 11 11:27:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	343831	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	511229	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	479437	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	154217	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	177676	52.04	ug/l	0.00
Spiked Amount 50.000			Recovery =	104.08%		
35) Dibromofluoromethane	7.59	113	157163	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery =	100.22%		
50) Toluene-d8	10.09	98	643826	53.34	ug/l	0.00
Spiked Amount 50.000			Recovery =	106.68%		
62) 4-Bromofluorobenzene	12.40	95	202174	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery =	98.52%		

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	2.18	62	3011685	805.551	ug/l	99
12) 1,1-Dichloroethene	3.72	96	125323	43.418	ug/l	94
21) trans-1,2-Dichloroethene	5.03	96	2529	0.818	ug/l	91
22) Diisopropyl ether	5.96	45	16699	1.772	ug/l	94
27) cis-1,2-Dichloroethene	6.83	96	338647	93.154	ug/l	89
39) Methylcyclohexane	9.08	83	26626	4.862	ug/l	99
40) Benzene	8.04	78	138135	10.346	ug/l	100
44) Trichloroethene	8.83	130	184253	48.027	ug/l	99
52) Toluene	10.16	92	3015	0.363	ug/l	99
65) Chlorobenzene	11.43	112	4964	0.520	ug/l	97
67) Ethyl Benzene	11.51	91	5105	0.308	ug/l	91
91) 1,2-Dichlorobenzene	13.65	146	6311	1.223	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN060719\
Data File : VN056051.D
Acq On : 6 Jun 2019 17:26
Operator : JC/SP
Sample : K3203-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LF015MW100-190604

Quant Time: Jun 11 11:27:57 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

