

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060719\
 Data File : VN056080.D
 Acq On : 7 Jun 2019 6:24
 Operator : JC/SP
 Sample : K3190-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BPS1-TT-MW304S-20190604

Quant Time: Jun 07 07:30:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	331671	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	502803	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	481298	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	151864	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.02	65	173032	52.54	ug/l	0.00
Spiked Amount						
			Recovery	=	105.08%	
35) Dibromofluoromethane	7.59	113	150623	48.83	ug/l	0.00
Spiked Amount						
			Recovery	=	97.66%	
50) Toluene-d8	10.09	98	634455	53.45	ug/l	0.00
Spiked Amount						
			Recovery	=	106.90%	
62) 4-Bromofluorobenzene	12.40	95	205003	50.78	ug/l	0.00
Spiked Amount						
			Recovery	=	101.56%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060719\
Data File : VN056080.D
Acq On : 7 Jun 2019 6:24
Operator : JC/SP
Sample : K3190-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPS1-TT-MW304S-20190604

Quant Time: Jun 07 07:30:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

