


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN060719\
Data File : VN056058.D
Acq On : 6 Jun 2019 20:47
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Title : SW846 8260
Last Update : Tue Jun 04 11:02:09 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	5290	PASS
75	95	30	60	46.2	15301	PASS
95	95	100	100	100.0	33098	PASS
96	95	5	9	6.3	2090	PASS
173	174	0.00	2	0.9	282	PASS
174	95	50	100	93.3	30866	PASS
175	174	5	9	7.6	2332	PASS
176	174	95	101	100.8	31122	PASS
177	176	5	9	6.5	2026	PASS