



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060722\
 Data File : VN072677.D
 Acq On : 07 Jun 2022 10:03
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N060622W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Mon Jun 06 14:10:15 2022



AutoFind: Averaged scan 1819 to 1839; Bkg corrected with scan 1818

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.5	3706	PASS
75	95	30	60	56.3	9266	PASS
95	95	100	100	100.0	16463	PASS
96	95	5	9	6.8	1119	PASS
173	174	0.00	2	1.0	102	PASS
174	95	50	100	62.8	10338	PASS
175	174	5	9	7.9	812	PASS
176	174	95	101	97.0	10027	PASS
177	176	5	9	6.3	632	PASS