

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060723\
 Data File : VN078095.D
 Acq On : 07 Jun 2023 11:52
 Operator : JC\MD
 Sample : 03039-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jun 07 16:40:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	395454	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	721580	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	617016	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	181261	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	264616	50.187	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.380%
35) Dibromofluoromethane	8.171	113	235284	51.570	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.140%
50) Toluene-d8	10.571	98	873306	49.651	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.300%
62) 4-Bromofluorobenzene	12.853	95	248430	45.434	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060723\
Data File : VN078095.D
Acq On : 07 Jun 2023 11:52
Operator : JC\MD
Sample : 03039-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Jun 07 16:40:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

