

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060723\
 Data File : VN078099.D
 Acq On : 07 Jun 2023 13:28
 Operator : JC\MD
 Sample : 03066-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 606

Quant Time: Jun 08 05:27:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	363063	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	674379	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	592051	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	189438	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	251521	51.959	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.920%	
35) Dibromofluoromethane	8.177	113	215481	50.535	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.060%	
50) Toluene-d8	10.571	98	821729	49.989	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.980%	
62) 4-Bromofluorobenzene	12.853	95	237069	46.391	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.780%	
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	9984	7.165	ug/l	97
17) Carbon Disulfide	4.724	76	5982	0.586	ug/l	100
25) 2-Butanone	7.500	43	18059	8.428	ug/l #	74
52) Toluene	10.641	92	3424	0.297	ug/l	96
80) 1,3,5-Trimethylbenzene	13.176	105	3881	0.268	ug/l	93
84) 1,2,4-Trimethylbenzene	13.488	105	7995	0.569	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060723\
Data File : VN078099.D
Acq On : 07 Jun 2023 13:28
Operator : JC\MD
Sample : 03066-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
606

Quant Time: Jun 08 05:27:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

